## An Example of the Confidence Limit

Present the results of following glucose analysis with the 95% confidence limit (CL).

Sample	% glucose
1	0.94
2	1.08
3	1.20
4	0.96
5	0.83
6	0.99

Average glucose = 1.00% standard deviation, s = 0.13